

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR082818\
 Data File : VR025648.D
 Acq On : 28 Aug 2018 15:06
 Operator : SY/MD
 Sample : VR0828WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK90

Quant Time: Aug 29 06:54:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 29 06:51:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	568996	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	467758	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	149234	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	164651	4.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.80%
7) Chloroethane-d5	2.63	69	138965	4.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	3.63	63	299212	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	6.59	46	171026	60.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.20%
24) Chloroform-d	7.20	84	289048	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	7.89	65	103630	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	7.86	84	722108	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	8.90	67	192519	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	9.97	98	669843	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	10.24	79	39985	4.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.00%
46) 2-Hexanone-d5	10.59	63	161465	53.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.28%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	77746	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	132617	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR082818\
Data File : VR025648.D
Acq On : 28 Aug 2018 15:06
Operator : SY/MD
Sample : VR0828WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK90

Quant Time: Aug 29 06:54:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 29 06:51:48 2018
Response via : Initial Calibration

