

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR082818\
 Data File : VR025652.D
 Acq On : 28 Aug 2018 18:29
 Operator : SY/MD
 Sample : J4601-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Aug 29 07:03:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 29 06:51:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	519593	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	434265	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	129837	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	158150	4.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.40%
7) Chloroethane-d5	2.64	69	130757	4.48	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	3.63	63	283612	3.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.80%
20) 2-Butanone-d5	6.59	46	150256	58.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.60%
24) Chloroform-d	7.21	84	259520	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	7.90	65	95253	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
32) Benzene-d6	7.86	84	631444	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
36) 1,2-Dichloropropane-d6	8.90	67	171205	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	9.97	98	582545	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	10.24	79	35427	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	10.59	63	153788	55.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.06%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	69648	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	114786	5.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR082818\
Data File : VR025652.D
Acq On : 28 Aug 2018 18:29
Operator : SY/MD
Sample : J4601-06
Misc : 25mL/MSVOA R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Aug 29 07:03:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 29 06:51:48 2018
Response via : Initial Calibration

