

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR082918\
 Data File : VR025656.D
 Acq On : 29 Aug 2018 13:57
 Operator : SY/MD
 Sample : VR0829WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK91

Quant Time: Aug 30 06:58:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 30 06:43:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	517485	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	428850	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	131843	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	142808	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	2.63	69	116258	4.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.00%
11) 1,1-Dichloroethene-d2	3.63	63	250659	3.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.20%
20) 2-Butanone-d5	6.60	46	127956	49.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.70%
24) Chloroform-d	7.20	84	230244	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.80%
26) 1,2-Dichloroethane-d4	7.90	65	84001	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
32) Benzene-d6	7.86	84	570475	4.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.40%
36) 1,2-Dichloropropane-d6	8.90	67	152196	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.20%
41) Toluene-d8	9.97	98	510876	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.60%
43) trans-1,3-Dichloropropene-	10.24	79	30124	3.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.40%
46) 2-Hexanone-d5	10.59	63	124326	45.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.10%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	61067	3.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	108608	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR082918\
Data File : VR025656.D
Acq On : 29 Aug 2018 13:57
Operator : SY/MD
Sample : VR0829WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK91

Quant Time: Aug 30 06:58:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 30 06:43:05 2018
Response via : Initial Calibration

