

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091818\
 Data File : VR025757.D
 Acq On : 18 Sep 2018 16:51
 Operator : SY/MD
 Sample : VR0918WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK98

Quant Time: Sep 19 09:22:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 16:10:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	778341	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	656924	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	192227	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	162389	3.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.40%
7) Chloroethane-d5	2.63	69	136130	3.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.60%
11) 1,1-Dichloroethene-d2	3.63	63	329149	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	6.59	46	161144	47.93	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.86%
24) Chloroform-d	7.20	84	352450	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.00%
26) 1,2-Dichloroethane-d4	7.90	65	129208	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
32) Benzene-d6	7.85	84	836100	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
36) 1,2-Dichloropropane-d6	8.90	67	232645	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.20%
41) Toluene-d8	9.97	98	761205	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.00%
43) trans-1,3-Dichloropropene-	10.24	79	44470	3.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.80%
46) 2-Hexanone-d5	10.59	63	175902	50.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.04%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	90689	3.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	70.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	150664	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR091818\
Data File : VR025757.D
Acq On : 18 Sep 2018 16:51
Operator : SY/MD
Sample : VR0918WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK98

Quant Time: Sep 19 09:22:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 16:10:44 2018
Response via : Initial Calibration

