

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
 Data File : VR025811.D
 Acq On : 21 Sep 2018 21:25
 Operator : SY/MD
 Sample : J4989-15
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ETOS6

Quant Time: Sep 22 05:30:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 22 04:37:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	465052	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	372099	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	126046	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	142866	4.29	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.80%
7) Chloroethane-d5	2.63	69	119395	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	3.62	63	256143	3.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.60%
20) 2-Butanone-d5	6.59	46	233825	52.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.34%
24) Chloroform-d	7.20	84	305000	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	7.90	65	138482	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	7.85	84	632710	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
36) 1,2-Dichloropropane-d6	8.90	67	176132	4.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.40%
41) Toluene-d8	9.97	98	558713	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	10.23	79	38485	3.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.80%
46) 2-Hexanone-d5	10.59	63	174814	50.21	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.42%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	67218	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	97585	4.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
Data File : VR025811.D
Acq On : 21 Sep 2018 21:25
Operator : SY/MD
Sample : J4989-15
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ETOS6

Quant Time: Sep 22 05:30:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 22 04:37:34 2018
Response via : Initial Calibration

