

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
 Data File : VR025814.D
 Acq On : 21 Sep 2018 23:04
 Operator : SY/MD
 Sample : J5030-01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEY70

Quant Time: Sep 22 05:40:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 22 04:37:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	450837	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	356236	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	117985	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	138208	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	2.62	69	112406	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.60%
11) 1,1-Dichloroethene-d2	3.61	63	237673	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	6.59	46	211802	49.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.44%
24) Chloroform-d	7.20	84	276629	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.60%
26) 1,2-Dichloroethane-d4	7.89	65	129603	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
32) Benzene-d6	7.85	84	573937	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
36) 1,2-Dichloropropane-d6	8.90	67	159328	4.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.40%
41) Toluene-d8	9.97	98	505026	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
43) trans-1,3-Dichloropropene-	10.24	79	33224	3.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.60%
46) 2-Hexanone-d5	10.59	63	151605	45.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.96%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60532	4.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	84471	4.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.80%

Target Compounds					Qvalue
47) Tetrachloroethene	10.51	164	10327	0.364 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
Data File : VR025814.D
Acq On : 21 Sep 2018 23:04
Operator : SY/MD
Sample : J5030-01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEY70

Quant Time: Sep 22 05:40:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 22 04:37:34 2018
Response via : Initial Calibration

