

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
 Data File : VR025805.D
 Acq On : 21 Sep 2018 17:33
 Operator : SY/MD
 Sample : VR0921WBL01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK82

Quant Time: Sep 22 05:09:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 22 04:37:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	461536	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	368505	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	123315	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	173477	5.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.00%
7) Chloroethane-d5	2.62	69	134066	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	3.60	63	302979	3.79	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.80%
20) 2-Butanone-d5	6.59	46	219767	49.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.76%
24) Chloroform-d	7.20	84	334748	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	7.90	65	150741	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	7.85	84	702954	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
36) 1,2-Dichloropropane-d6	8.90	67	188873	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	9.97	98	622249	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
43) trans-1,3-Dichloropropene-	10.24	79	48097	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	10.59	63	174254	50.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.08%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	70496	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	107620	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
Data File : VR025805.D
Acq On : 21 Sep 2018 17:33
Operator : SY/MD
Sample : VR0921WBL01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK82

Quant Time: Sep 22 05:09:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 22 04:37:34 2018
Response via : Initial Calibration

