

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
 Data File : VR025812.D
 Acq On : 21 Sep 2018 21:58
 Operator : SY/MD
 Sample : J4989-16
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ET0S7

Quant Time: Sep 22 05:34:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 22 04:37:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	452959	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	368830	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	117521	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	149983	4.62	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	2.63	69	117900	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	3.61	63	256065	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	6.59	46	232581	53.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.58%
24) Chloroform-d	7.20	84	297654	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	7.90	65	136044	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	7.85	84	618759	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	8.89	67	174409	4.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.40%
41) Toluene-d8	9.97	98	545007	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	10.24	79	36813	3.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.20%
46) 2-Hexanone-d5	10.59	63	165606	47.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.98%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	63925	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	94278	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%


Target Compounds					Qvalue
13) Acetone	3.70	43	56241	19.712 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092118\
Data File : VR025812.D
Acq On : 21 Sep 2018 21:58
Operator : SY/MD
Sample : J4989-16
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ET0S7

Quant Time: Sep 22 05:34:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 22 04:37:34 2018
Response via : Initial Calibration





#13

Acetone

Concen: 19.712 ug/L

RT: 3.70 min Scan# 332

Delta R.T. 0.01 min

Lab File: VR025812.D

Acq: 21 Sep 2018 21:58

Instrument :
MSVOA_R
ClientSampleId :
ET0S7

Tgt Ion: 43 Resp: 56241

Ion Ratio Lower Upper

43 100

58 26.1 0.0 51.4

