

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
 Data File : VR025883.D
 Acq On : 27 Sep 2018 20:35
 Operator : SY/MD
 Sample : J5119-04
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 COL04

Quant Time: Sep 28 06:53:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 05:44:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	489182	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	403350	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	134342	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	197113	5.11	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.20%
7) Chloroethane-d5	2.62	69	168126	5.33	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	3.63	63	344820	3.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.20%
20) 2-Butanone-d5	6.59	46	236681	55.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.08%
24) Chloroform-d	7.20	84	319714	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	7.89	65	145208	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	7.85	84	671221	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	8.90	67	182670	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	9.97	98	618038	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
43) trans-1,3-Dichloropropene-	10.24	79	36928	4.50	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.00%
46) 2-Hexanone-d5	10.59	63	172600	55.12	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.24%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	70311	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	108820	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
Data File : VR025883.D
Acq On : 27 Sep 2018 20:35
Operator : SY/MD
Sample : J5119-04
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
COL04

Quant Time: Sep 28 06:53:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 05:44:06 2018
Response via : Initial Calibration

