

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
 Data File : VR025884.D
 Acq On : 27 Sep 2018 21:08
 Operator : SY/MD
 Sample : J5119-05
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 COL05

Quant Time: Sep 28 06:56:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 05:44:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	506706	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	417489	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	139605	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	188371	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	2.63	69	160963	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.60%
11) 1,1-Dichloroethene-d2	3.63	63	340225	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	6.59	46	232420	52.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.36%
24) Chloroform-d	7.20	84	313123	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.00%
26) 1,2-Dichloroethane-d4	7.90	65	140759	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
32) Benzene-d6	7.85	84	668778	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	8.90	67	181341	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	9.97	98	610021	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
43) trans-1,3-Dichloropropene-	10.23	79	33583	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	10.59	63	165629	51.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.20%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	71808	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	106852	4.78	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.60%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
Data File : VR025884.D
Acq On : 27 Sep 2018 21:08
Operator : SY/MD
Sample : J5119-05
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
COL05

Quant Time: Sep 28 06:56:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 05:44:06 2018
Response via : Initial Calibration

