

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
 Data File : VR025886.D
 Acq On : 27 Sep 2018 22:14
 Operator : SY/MD
 Sample : J5119-07
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 COL07

Quant Time: Sep 28 07:02:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 05:44:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	510524	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	413801	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	144126	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	190742	4.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.80%
7) Chloroethane-d5	2.63	69	154074	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	3.64	63	325889	3.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.00%
20) 2-Butanone-d5	6.59	46	212617	47.37	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.74%
24) Chloroform-d	7.20	84	297593	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	7.90	65	131962	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
32) Benzene-d6	7.85	84	625558	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
36) 1,2-Dichloropropane-d6	8.90	67	166018	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	9.97	98	566763	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
43) trans-1,3-Dichloropropene-	10.24	79	32953	3.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.20%
46) 2-Hexanone-d5	10.59	63	159390	49.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	66513	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	100504	4.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.00%

Target Compounds					Qvalue
34) Trichloroethene	8.71	95	47990	1.312 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
Data File : VR025886.D
Acq On : 27 Sep 2018 22:14
Operator : SY/MD
Sample : J5119-07
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0L07

Quant Time: Sep 28 07:02:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 05:44:06 2018
Response via : Initial Calibration

