

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
 Data File : VR025880.D
 Acq On : 27 Sep 2018 18:56
 Operator : SY/MD
 Sample : J5119-01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 COL01

Quant Time: Sep 28 06:47:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 05:44:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	533562	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	426385	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	143268	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	198918	4.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.60%
7) Chloroethane-d5	2.63	69	158978	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.40%
11) 1,1-Dichloroethene-d2	3.63	63	345324	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	6.59	46	236612	50.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.88%
24) Chloroform-d	7.20	84	328580	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
26) 1,2-Dichloroethane-d4	7.90	65	147203	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	7.85	84	692056	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	8.90	67	184194	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	9.97	98	623180	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
43) trans-1,3-Dichloropropene-	10.23	79	37650	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	10.59	63	172233	52.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.06%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	71013	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	109245	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092718\
Data File : VR025880.D
Acq On : 27 Sep 2018 18:56
Operator : SY/MD
Sample : J5119-01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
COL01

Quant Time: Sep 28 06:47:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 05:44:06 2018
Response via : Initial Calibration

