

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092818\
 Data File : VR025904.D
 Acq On : 28 Sep 2018 16:57
 Operator : SY/MD
 Sample : J5119-11
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Sep 29 05:53:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 29 05:14:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	511638	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	422678	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	139583	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	184664	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	2.63	69	157245	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.40%
11) 1,1-Dichloroethene-d2	3.63	63	342088	3.61	ug/L	0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	6.58	46	226204	50.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.58%
24) Chloroform-d	7.20	84	310961	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
26) 1,2-Dichloroethane-d4	7.89	65	140630	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	7.85	84	642897	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	8.90	67	171132	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.20%
41) Toluene-d8	9.97	98	584212	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	10.24	79	35821	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	10.59	63	160723	48.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.96%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64047	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	105422	4.72	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092818\
Data File : VR025904.D
Acq On : 28 Sep 2018 16:57
Operator : SY/MD
Sample : J5119-11
Misc : 25mL/MSVOA R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Sep 29 05:53:06 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 29 05:14:56 2018
Response via : Initial Calibration

