

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025951.D
 Acq On : 5 Oct 2018 11:28
 Operator : SY/MD
 Sample : VR1005WBL01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK11

Quant Time: Oct 06 04:31:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:21:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	441865	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	349895	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	115430	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	199678	5.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.60%
7) Chloroethane-d5	2.63	69	169355	5.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.00%
11) 1,1-Dichloroethene-d2	3.63	63	369686	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.40%
20) 2-Butanone-d5	6.59	46	184345	47.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.92%
24) Chloroform-d	7.20	84	293874	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	7.90	65	130875	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	7.85	84	611064	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	8.90	67	158046	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	9.97	98	575494	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
43) trans-1,3-Dichloropropene-	10.24	79	37714	5.30	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.00%
46) 2-Hexanone-d5	10.59	63	129982	47.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.70%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	59306	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	99019	5.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
Data File : VR025951.D
Acq On : 5 Oct 2018 11:28
Operator : SY/MD
Sample : VR1005WBL01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK11

Quant Time: Oct 06 04:31:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 04:21:51 2018
Response via : Initial Calibration

