

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025952.D
 Acq On : 5 Oct 2018 12:06
 Operator : SY/MD
 Sample : J5258-01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ESNN1

Quant Time: Oct 06 04:42:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:21:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	452557	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	375759	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	130775	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	129843	3.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.80%
7) Chloroethane-d5	2.63	69	132094	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.60%
11) 1,1-Dichloroethene-d2	3.63	63	246562	2.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.80%#
20) 2-Butanone-d5	6.59	46	203898	51.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.50%
24) Chloroform-d	7.20	84	251181	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
26) 1,2-Dichloroethane-d4	7.90	65	115281	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
32) Benzene-d6	7.85	84	463035	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.00%
36) 1,2-Dichloropropane-d6	8.90	67	134170	3.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.60%
41) Toluene-d8	9.97	98	400879	3.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.20%#
43) trans-1,3-Dichloropropene-	10.24	79	30693	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
46) 2-Hexanone-d5	10.59	63	138968	47.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.28%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	59231	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	85717	4.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.80%

Target Compounds					Qvalue
13) Acetone	3.70	43	30438	10.555 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
Data File : VR025952.D
Acq On : 5 Oct 2018 12:06
Operator : SY/MD
Sample : J5258-01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ESNN1

Quant Time: Oct 06 04:42:45 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 04:21:51 2018
Response via : Initial Calibration

