

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025953.D
 Acq On : 5 Oct 2018 12:42
 Operator : SY/MD
 Sample : J5258-06
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Oct 06 04:47:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:21:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	422179	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	355117	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	116897	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	190029	5.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.20%
7) Chloroethane-d5	2.63	69	166152	6.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.20%
11) 1,1-Dichloroethene-d2	3.63	63	344978	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.20%
20) 2-Butanone-d5	6.59	46	186323	50.20	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.40%
24) Chloroform-d	7.20	84	281587	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	7.90	65	126981	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	7.85	84	576803	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	8.90	67	157168	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	9.97	98	549326	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
43) trans-1,3-Dichloropropene-	10.24	79	37601	5.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.00%
46) 2-Hexanone-d5	10.59	63	137456	49.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.72%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	58749	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	96865	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
Data File : VR025953.D
Acq On : 5 Oct 2018 12:42
Operator : SY/MD
Sample : J5258-06
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Oct 06 04:47:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 04:21:51 2018
Response via : Initial Calibration

