

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025962.D
 Acq On : 5 Oct 2018 18:46
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_R
 LabSampleId :
 VSTD00582

Quant Time: Oct 06 05:24:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:21:51 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	5.000	4.000	20.0	64	0.00
3 T	Chloromethane	5.000	4.849	3.0	84	0.00
4 S	Vinyl Chloride-d3	5.000	5.281	-5.6	84	0.00
5 T	Vinyl chloride	5.000	4.850	3.0	82	0.00
6 T	Bromomethane	5.000	4.811	3.8	83	0.00
7 S	Chloroethane-d5	5.000	5.577	-11.5	91	0.00
8 T	Chloroethane	5.000	4.915	1.7	83	0.00
9 T	Trichlorofluoromethane	5.000	5.200	-4.0	85	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.081	-1.6	82	0.00
11 S	1,1-Dichloroethene-d2	5.000	5.148	-3.0	84	0.00
12 T	1,1-Dichloroethene	5.000	5.043	-0.9	85	0.00
13 T	Acetone	50.000	54.970	-9.9	98	0.00
14 T	Carbon disulfide	5.000	4.407	11.9	71	0.00
15 T	Methyl Acetate	5.000	5.017	-0.3	82	0.00
16 T	Methylene chloride	5.000	4.953	0.9	89	0.00
17 T	Methyl tert-butyl Ether	5.000	4.667	6.7	79	0.00
18 T	trans-1,2-Dichloroethene	5.000	4.436	11.3	74	0.00
19 T	1,1-Dichloroethane	5.000	4.487	10.3	76	0.00
20 S	2-Butanone-d5	50.000	52.008	-4.0	81	0.00
21 T	2-Butanone	50.000	48.361	3.3	81	0.00
22 T	cis-1,2-Dichloroethene	5.000	4.516	9.7	77	0.00
23 T	Bromochloromethane	5.000	4.724	5.5	79	0.00
24 S	Chloroform-d	5.000	5.058	-1.2	83	0.00
25 T	Chloroform	5.000	4.627	7.5	79	0.00
26 S	1,2-Dichloroethane-d4	5.000	5.205	-4.1	86	0.00
27 T	1,2-Dichloroethane	5.000	4.667	6.7	77	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	83	0.00
29 T	1,1,1-Trichloroethane	5.000	4.526	9.5	78	0.00
30 T	Cyclohexane	5.000	3.966	20.7	66	0.00
31 T	Carbon tetrachloride	5.000	4.621	7.6	80	0.00
32 S	Benzene-d6	5.000	4.607	7.9	77	0.00
33 T	Benzene	5.000	4.386	12.3	78	0.00
34 T	Trichloroethene	5.000	4.222	15.6	76	0.00
35 T	Methylcyclohexane	5.000	4.104	17.9	67	0.00
36 S	1,2-Dichloropropane-d6	5.000	4.604	7.9	77	0.00
37 T	1,2-Dichloropropane	5.000	4.432	11.4	76	0.00
38 T	Bromodichloromethane	5.000	4.562	8.8	78	0.00
39 T	cis-1,3-Dichloropropene	5.000	4.864	2.7	78	0.00
40 T	4-Methyl-2-pentanone	50.000	48.180	3.6	81	0.00
41 S	Toluene-d8	5.000	4.748	5.0	79	0.00
42 T	Toluene	5.000	4.472	10.6	77	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	5.377	-7.5	91	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.203	-4.1	87	0.00
45 T	1,1,2-Trichloroethane	5.000	4.344	13.1	77	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025962.D
 Acq On : 5 Oct 2018 18:46
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_R
 LabSampleId :
 VSTD00582

Quant Time: Oct 06 05:24:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 04:21:51 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	52.684	-5.4	83	0.00
47 T	Tetrachloroethene	5.000	4.357	12.9	75	0.00
48 T	2-Hexanone	50.000	49.151	1.7	83	0.00
49 T	Dibromochloromethane	5.000	4.763	4.7	83	0.00
50 T	1,2-Dibromoethane	5.000	4.641	7.2	78	0.00
51 T	Chlorobenzene	5.000	4.585	8.3	81	0.00
52 T	Ethylbenzene	5.000	4.617	7.7	79	0.00
53 T	m,p-Xylene	5.000	4.625	7.5	79	0.00
54 T	o-Xylene	5.000	4.609	7.8	80	0.00
55 T	Styrene	5.000	4.778	4.4	81	0.00
56 T	Isopropylbenzene	5.000	4.728	5.4	80	0.00
57 S	1,1,2,2-Tetrachloroethane-d	5.000	4.847	3.1	79	0.00
58 T	1,1,2,2-Tetrachloroethane	5.000	4.611	7.8	80	0.00
59	1,2,3-Trichloropropane	5.000	4.753	4.9	85	0.00
60 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	88	0.00
61 T	Bromoform	5.000	4.681	6.4	84	0.00
62 T	1,3-Dichlorobenzene	5.000	4.400	12.0	79	0.00
63 T	1,4-Dichlorobenzene	5.000	4.513	9.7	83	0.00
64 S	1,2-Dichlorobenzene-d4	5.000	4.717	5.7	80	0.00
65 T	1,2-Dichlorobenzene	5.000	4.513	9.7	83	0.00
66 T	1,2-Dibromo-3-chloropropane	5.000	4.040	19.2	95	0.00
67	1,3,5-Trichlorobenzene	5.000	4.556	8.9	82	0.00
68 T	1,2,4-trichlorobenzene	5.000	4.581	8.4	87	0.00
69	Naphthalene	5.000	4.110	17.8	72	0.00
70 T	1,2,3-Trichlorobenzene	5.000	4.891	2.2	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0