

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025966.D
 Acq On : 5 Oct 2018 22:03
 Operator : SY/MD
 Sample : J5276-11
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYA8

Quant Time: Oct 06 05:45:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 05:30:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	402193	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	340791	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	112080	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	109631	3.46	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.20%
7) Chloroethane-d5	2.63	69	111935	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.40%
11) 1,1-Dichloroethene-d2	3.63	63	205590	2.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.20%#
20) 2-Butanone-d5	6.59	46	189831	53.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.38%
24) Chloroform-d	7.20	84	222484	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.60%
26) 1,2-Dichloroethane-d4	7.90	65	105065	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
32) Benzene-d6	7.85	84	394094	3.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	69.60%#
36) 1,2-Dichloropropane-d6	8.89	67	122663	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.20%
41) Toluene-d8	9.97	98	335199	3.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.00%#
43) trans-1,3-Dichloropropene-	10.23	79	25684	3.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.00%
46) 2-Hexanone-d5	10.59	63	138250	52.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.50%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	52124	4.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	74628	4.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
Data File : VR025966.D
Acq On : 5 Oct 2018 22:03
Operator : SY/MD
Sample : J5276-11
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYA8

Quant Time: Oct 06 05:45:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 05:30:02 2018
Response via : Initial Calibration

