

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
 Data File : VR025972.D
 Acq On : 6 Oct 2018 1:21
 Operator : SY/MD
 Sample : J5276-17
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYB4

Quant Time: Oct 06 06:27:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 06 05:30:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	425182	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	351288	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	112116	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	108309	3.23	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.60%
7) Chloroethane-d5	2.63	69	111081	4.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.00%
11) 1,1-Dichloroethene-d2	3.63	63	200339	2.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.80%#
20) 2-Butanone-d5	6.59	46	197957	52.96	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.92%
24) Chloroform-d	7.20	84	224381	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.80%
26) 1,2-Dichloroethane-d4	7.90	65	104917	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
32) Benzene-d6	7.85	84	388689	3.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.60%#
36) 1,2-Dichloropropane-d6	8.89	67	118190	3.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.00%
41) Toluene-d8	9.97	98	323909	2.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	59.00%#
43) trans-1,3-Dichloropropene-	10.24	79	22187	3.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.00%
46) 2-Hexanone-d5	10.59	63	142259	52.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.32%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	54181	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	71605	3.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	79.80%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR100518\
Data File : VR025972.D
Acq On : 6 Oct 2018 1:21
Operator : SY/MD
Sample : J5276-17
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYB4

Quant Time: Oct 06 06:27:06 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 06 05:30:02 2018
Response via : Initial Calibration

