

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR025993.D
 Acq On : 8 Oct 2018 18:58
 Operator : SY/MD
 Sample : J5279-07
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYG4

Quant Time: Oct 09 06:39:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 05:29:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	429472	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	360285	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	116700	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	173498	5.12	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	2.63	69	157783	5.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.00%
11) 1,1-Dichloroethene-d2	3.63	63	310964	3.91	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.20%
20) 2-Butanone-d5	6.59	46	201696	53.42	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.84%
24) Chloroform-d	7.20	84	282719	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	7.89	65	132241	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	7.85	84	553069	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
36) 1,2-Dichloropropane-d6	8.90	67	151789	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	9.97	98	510752	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	10.24	79	35328	4.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.40%
46) 2-Hexanone-d5	10.59	63	158354	56.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65671	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	99503	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds					Ovalue
13) Acetone	3.70	43	9443	3.451 ug/L	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR025993.D
Acq On : 8 Oct 2018 18:58
Operator : SY/MD
Sample : J5279-07
Misc : 25mL/MSVOA R/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYG4

Quant Time: Oct 09 06:39:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 05:29:53 2018
Response via : Initial Calibration

