

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR026001.D
 Acq On : 9 Oct 2018 1:31
 Operator : SY/MD
 Sample : J5279-13
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYH0

Quant Time: Oct 09 07:10:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 06:51:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	410076	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	340750	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	110054	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	174912	5.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.20%
7) Chloroethane-d5	2.62	69	158975	6.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.40%
11) 1,1-Dichloroethene-d2	3.62	63	309768	4.08	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.60%
20) 2-Butanone-d5	6.59	46	191254	53.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.10%
24) Chloroform-d	7.20	84	272757	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	7.90	65	122451	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	7.85	84	530758	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	8.90	67	150512	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	9.97	98	493413	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
43) trans-1,3-Dichloropropene-	10.24	79	30859	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	10.59	63	142904	54.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.04%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60745	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	96649	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR026001.D
Acq On : 9 Oct 2018 1:31
Operator : SY/MD
Sample : J5279-13
Misc : 25mL/MSVOA R/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYH0

Quant Time: Oct 09 07:10:08 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 06:51:21 2018
Response via : Initial Calibration

