

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR026004.D
 Acq On : 9 Oct 2018 3:09
 Operator : SY/MD
 Sample : J5279-16
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYH3

Quant Time: Oct 09 07:18:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 06:51:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	415228	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	355533	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	112269	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	183056	5.59	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.80%
7) Chloroethane-d5	2.63	69	161684	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.80%
11) 1,1-Dichloroethene-d2	3.61	63	330756	4.30	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.00%
20) 2-Butanone-d5	6.59	46	197543	54.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.24%
24) Chloroform-d	7.20	84	281231	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	7.89	65	130936	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
32) Benzene-d6	7.85	84	566214	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	8.90	67	154852	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	9.97	98	522943	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	10.24	79	31655	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	10.59	63	148818	53.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.82%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64270	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	98722	5.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.80%

Target Compounds					Ovalue
13) Acetone	3.70	43	7367	2.784 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR026004.D
Acq On : 9 Oct 2018 3:09
Operator : SY/MD
Sample : J5279-16
Misc : 25mL/MSVOA R/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYH3

Quant Time: Oct 09 07:18:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 06:51:21 2018
Response via : Initial Calibration

