

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR025980.D
 Acq On : 8 Oct 2018 11:32
 Operator : SY/MD
 Sample : VR1008WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK13

Quant Time: Oct 09 05:39:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 05:29:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	520490	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	430242	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	149160	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.17	65	175995	4.29	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.80%
7) Chloroethane-d5	2.63	69	157403	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	3.63	63	321652	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	6.59	46	196894	43.03	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.06%
24) Chloroform-d	7.20	84	302469	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	7.90	65	136613	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	7.85	84	605690	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
36) 1,2-Dichloropropane-d6	8.90	67	162421	4.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.20%
41) Toluene-d8	9.97	98	553259	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	10.24	79	40500	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
46) 2-Hexanone-d5	10.59	63	145191	43.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.94%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	61991	3.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	108824	4.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR025980.D
Acq On : 8 Oct 2018 11:32
Operator : SY/MD
Sample : VR1008WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK13

Quant Time: Oct 09 05:39:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 05:29:53 2018
Response via : Initial Calibration

