

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR025988.D
 Acq On : 8 Oct 2018 16:13
 Operator : SY/MD
 Sample : J5276-18
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYB5

Quant Time: Oct 09 06:28:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 05:29:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	395461	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	337092	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	119271	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	180807	5.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.00%
7) Chloroethane-d5	2.63	69	165109	6.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	129.60%
11) 1,1-Dichloroethene-d2	3.63	63	324998	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.80%
20) 2-Butanone-d5	6.59	46	166914	48.01	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.02%
24) Chloroform-d	7.20	84	279062	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	7.90	65	125089	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	7.85	84	552557	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	8.90	67	155299	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	9.97	98	506048	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
43) trans-1,3-Dichloropropene-	10.24	79	34228	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	10.59	63	149571	57.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.30%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64636	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	102184	5.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR025988.D
Acq On : 8 Oct 2018 16:13
Operator : SY/MD
Sample : J5276-18
Misc : 25mL/MSVOA R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYB5

Quant Time: Oct 09 06:28:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 05:29:53 2018
Response via : Initial Calibration

