

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
 Data File : VR025998.D
 Acq On : 8 Oct 2018 23:53
 Operator : SY/MD
 Sample : J5279-10
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYG7

Quant Time: Oct 09 07:03:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 09 06:51:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	417741	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	354673	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	115517	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	171332	5.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.00%
7) Chloroethane-d5	2.63	69	151036	5.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.20%
11) 1,1-Dichloroethene-d2	3.63	63	312101	4.03	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.60%
20) 2-Butanone-d5	6.58	46	199268	54.26	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.52%
24) Chloroform-d	7.20	84	276744	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	7.90	65	127181	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	7.85	84	540763	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
36) 1,2-Dichloropropane-d6	8.89	67	150613	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	9.97	98	494483	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
43) trans-1,3-Dichloropropene-	10.23	79	32512	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	10.59	63	149503	54.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.58%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60241	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	95218	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100818\
Data File : VR025998.D
Acq On : 8 Oct 2018 23:53
Operator : SY/MD
Sample : J5279-10
Misc : 25mL/MSVOA R/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYG7

Quant Time: Oct 09 07:03:43 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 09 06:51:21 2018
Response via : Initial Calibration

