

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101018\
 Data File : VR026024.D
 Acq On : 10 Oct 2018 11:58
 Operator : SY/MD
 Sample : J5276-13DL 4X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEYB0DL

Quant Time: Oct 11 04:38:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 11 04:11:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	489057	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	407665	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	136728	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	170441	4.42	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	2.62	69	151193	4.80	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	3.63	63	322565	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	6.59	46	233357	54.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.56%
24) Chloroform-d	7.20	84	313456	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	7.89	65	143181	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	7.85	84	629196	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
36) 1,2-Dichloropropane-d6	8.90	67	174471	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	9.97	98	569703	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	10.24	79	32700	3.94	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.80%
46) 2-Hexanone-d5	10.59	63	174769	55.22	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.44%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	69243	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	110260	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
13) Acetone	3.70	43	5867	1.883 ug/L	73
29) 1,1,1-Trichloroethane	7.44	97	5750	0.125 ug/L	97
34) Trichloroethene	8.71	95	204232	5.669 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101018\
Data File : VR026024.D
Acq On : 10 Oct 2018 11:58
Operator : SY/MD
Sample : J5276-13DL 4X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEYB0DL

Quant Time: Oct 11 04:38:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 11 04:11:41 2018
Response via : Initial Calibration

