

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110518\
 Data File : VR026178.D
 Acq On : 5 Nov 2018 14:09
 Operator : SY/MD
 Sample : J5802-12
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0BD8

Quant Time: Nov 06 03:38:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 03:06:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	548019	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	451029	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	137383	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	223578	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	2.62	69	208103	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	3.60	63	448617	3.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.20%
20) 2-Butanone-d5	6.58	46	196452	49.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.78%
24) Chloroform-d	7.20	84	377720	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	7.89	65	158066	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	7.85	84	761406	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	8.89	67	208261	5.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.20%
41) Toluene-d8	9.97	98	736956	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
43) trans-1,3-Dichloropropene-	10.24	79	48871	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	10.59	63	157750	55.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.32%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	75201	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	129225	5.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR110518\
Data File : VR026178.D
Acq On : 5 Nov 2018 14:09
Operator : SY/MD
Sample : J5802-12
Misc : 25mL/MSVOA R/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0BD8

Quant Time: Nov 06 03:38:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 03:06:52 2018
Response via : Initial Calibration

