

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110818\
 Data File : VR026211.D
 Acq On : 8 Nov 2018 12:22
 Operator : SY/MD
 Sample : J5802-08
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Nov 12 06:59:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 15:50:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	558628	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	463986	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	138116	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	244223	5.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.00%
7) Chloroethane-d5	2.63	69	224602	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.80%
11) 1,1-Dichloroethene-d2	3.60	63	492166	4.10	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.00%
20) 2-Butanone-d5	6.58	46	205177	51.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.24%
24) Chloroform-d	7.20	84	395568	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
26) 1,2-Dichloroethane-d4	7.90	65	159053	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	7.85	84	807483	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
36) 1,2-Dichloropropane-d6	8.89	67	214812	5.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.60%
41) Toluene-d8	9.97	98	743187	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
43) trans-1,3-Dichloropropene-	10.24	79	47409	4.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.00%
46) 2-Hexanone-d5	10.59	63	151900	52.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.20%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	71811	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	126892	5.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110818\
Data File : VR026211.D
Acq On : 8 Nov 2018 12:22
Operator : SY/MD
Sample : J5802-08
Misc : 25mL/MSVOA R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Nov 12 06:59:09 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 15:50:53 2018
Response via : Initial Calibration

