

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR122118\
 Data File : VR026290.D
 Acq On : 21 Dec 2018 20:05
 Operator : SY/MD
 Sample : J6535-06
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0TR1

Quant Time: Dec 21 22:52:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 21 22:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	506244	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	425429	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	119863	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	142287	4.42	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	2.63	69	150842	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	3.61	63	311089	3.42	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	6.59	46	155567	54.38	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.76%
24) Chloroform-d	7.21	84	328184	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	7.90	65	131513	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	7.86	84	612536	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
36) 1,2-Dichloropropane-d6	8.90	67	165809	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	9.97	98	557965	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	10.24	79	36668	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
46) 2-Hexanone-d5	10.59	63	103916	54.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.26%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64402	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	111486	5.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR122118\
Data File : VR026290.D
Acq On : 21 Dec 2018 20:05
Operator : SY/MD
Sample : J6535-06
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0TR1

Quant Time: Dec 21 22:52:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 21 22:24:46 2018
Response via : Initial Calibration

