

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR122118\
 Data File : VR026292.D
 Acq On : 21 Dec 2018 21:11
 Operator : SY/MD
 Sample : J6536-02
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BE809

Quant Time: Dec 21 22:59:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 21 22:24:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	490286	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	455689	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	209660	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	145619	4.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.40%
7) Chloroethane-d5	2.63	69	160358	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.60%
11) 1,1-Dichloroethene-d2	3.61	63	314193	3.56	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	6.59	46	119788	43.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.46%
24) Chloroform-d	7.20	84	325478	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	7.90	65	131597	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
32) Benzene-d6	7.85	84	643221	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	8.90	67	157438	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.60%
41) Toluene-d8	9.97	98	575786	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.80%
43) trans-1,3-Dichloropropene-	10.24	79	41229	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	10.59	63	129666	63.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	127.30%
57) 1,1,2,2-Tetrachloroethane-	12.37	84	79683	5.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	171792	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Chloroethane	2.66	64	23355	0.767	ug/L	98
19) 1,1-Dichloroethane	5.69	63	86264	1.213	ug/L	96
30) Cyclohexane	7.52	56	103963	1.572	ug/L	99
33) Benzene	7.91	78	1164138	6.490	ug/L	100
35) Methylcyclohexane	8.96	83	42578	0.621	ug/L	89
42) Toluene	10.04	91	25268	0.138	ug/L	95
51) Chlorobenzene	11.32	112	29788	0.290	ug/L	98
52) Ethylbenzene	11.39	91	67308	0.325	ug/L	95
56) Isopropylbenzene	12.13	105	10237523	53.382	ug/L #	81
65) 1,2-Dichlorobenzene	13.53	146	21488	0.387	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR122118\
Data File : VR026292.D
Acq On : 21 Dec 2018 21:11
Operator : SY/MD
Sample : J6536-02
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BE809

Quant Time: Dec 21 22:59:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 21 22:24:46 2018
Response via : Initial Calibration

