

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010224\
 Data File : VY016894.D
 Acq On : 02 Jan 2024 14:21
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY010224

Quant Time: Jan 03 00:17:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
 Supervised By :Semsettin Yesilyurt 01/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	150756	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	250199	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	212664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	95726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	67239	45.126	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	90.260%		
35) Dibromofluoromethane	7.721	113	75114	47.112	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.220%		
50) Toluene-d8	10.191	98	264060	48.097	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	96.200%		
62) 4-Bromofluorobenzene	12.489	95	87401	47.184	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	94.360%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.905	85	54432	47.846	ug/l	97
3) Chloromethane	2.119	50	85229	48.126	ug/l	99
4) Vinyl Chloride	2.253	62	101834	48.749	ug/l	98
5) Bromomethane	2.661	94	72921	48.727	ug/l	94
6) Chloroethane	2.802	64	68326	48.353	ug/l	99
7) Trichlorofluoromethane	3.131	101	125812	48.881	ug/l	96
8) Diethyl Ether	3.533	74	39150	47.596	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.911	101	80655	48.194	ug/l	100
10) Methyl Iodide	4.100	142	73800	51.835	ug/l	99
11) Tert butyl alcohol	4.954	59	24419	221.493	ug/l	99
12) 1,1-Dichloroethene	3.881	96	71569	49.843	ug/l	94
13) Acrolein	3.734	56	35924	220.149	ug/l	98
14) Allyl chloride	4.484	41	115466	50.866	ug/l	100
15) Acrylonitrile	5.167	53	87503	227.431	ug/l	99
16) Acetone	3.954	43	83103	223.763	ug/l	100
17) Carbon Disulfide	4.204	76	174734	50.929	ug/l	99
18) Methyl Acetate	4.484	43	79697	49.924	ug/l	99
19) Methyl tert-butyl Ether	5.228	73	181651	48.626	ug/l	98
20) Methylene Chloride	4.722	84	99859	52.404	ug/l	98
21) trans-1,2-Dichloroethene	5.228	96	82436	50.127	ug/l	99
22) Diisopropyl ether	6.130	45	268198	51.105	ug/l	94
23) Vinyl Acetate	6.069	43	573952	238.518	ug/l	100
24) 1,1-Dichloroethane	6.027	63	155221	49.219	ug/l	98
25) 2-Butanone	6.990	43	111728	218.509	ug/l	100
26) 2,2-Dichloropropane	6.990	77	137058	50.124	ug/l	100
27) cis-1,2-Dichloroethene	6.996	96	97058	50.347	ug/l	98
28) Bromochloromethane	7.337	49	63151	53.339	ug/l	99
29) Tetrahydrofuran	7.356	42	66609	220.872	ug/l	99
30) Chloroform	7.514	83	160332	49.334	ug/l	100
31) Cyclohexane	7.795	56	125477	49.065	ug/l	99
32) 1,1,1-Trichloroethane	7.709	97	138395	50.223	ug/l	100
36) 1,1-Dichloropropene	7.929	75	118548	51.020	ug/l	100
37) Ethyl Acetate	7.081	43	46597	44.129	ug/l	100
38) Carbon Tetrachloride	7.910	117	119299	50.535	ug/l	100
39) Methylcyclohexane	9.191	83	138463	53.008	ug/l	99
40) Benzene	8.167	78	350830	50.333	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010224\
 Data File : VY016894.D
 Acq On : 02 Jan 2024 14:21
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY010224

Quant Time: Jan 03 00:17:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
 Supervised By :Semsettin Yesilyurt 01/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	28007	50.362	ug/l	97
42) 1,2-Dichloroethane	8.246	62	88992	48.093	ug/l	99
43) Isopropyl Acetate	8.282	43	92773	45.880	ug/l	100
44) Trichloroethene	8.947	130	90148	50.827	ug/l	96
45) 1,2-Dichloropropane	9.227	63	89277	49.718	ug/l	99
46) Dibromomethane	9.313	93	44332	48.408	ug/l	98
47) Bromodichloromethane	9.508	83	120872	49.151	ug/l	97
48) Methyl methacrylate	9.300	41	53814	49.314	ug/l	99
49) 1,4-Dioxane	9.307	88	8808	928.874	ug/l	97
51) 4-Methyl-2-Pentanone	10.075	43	246177	228.923	ug/l	99
52) Toluene	10.252	92	214834	51.072	ug/l	98
53) t-1,3-Dichloropropene	10.471	75	110224	50.038	ug/l	99
54) cis-1,3-Dichloropropene	9.935	75	134005	50.656	ug/l	96
55) 1,1,2-Trichloroethane	10.654	97	63464	47.435	ug/l	96
56) Ethyl methacrylate	10.520	69	55715	49.165	ug/l	98
57) 1,3-Dichloropropane	10.800	76	109039	48.814	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.788	63	167516	235.967	ug/l	99
59) 2-Hexanone	10.843	43	172631	218.016	ug/l	100
60) Dibromochloromethane	10.995	129	75242	48.766	ug/l	99
61) 1,2-Dibromoethane	11.099	107	57180	49.190	ug/l	97
64) Tetrachloroethene	10.727	164	69275	50.245	ug/l	98
65) Chlorobenzene	11.526	112	226858	51.207	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.599	131	85180	50.579	ug/l	99
67) Ethyl Benzene	11.599	91	421349	52.828	ug/l	98
68) m/p-Xylenes	11.709	106	311476	106.392	ug/l	99
69) o-Xylene	12.038	106	145587	53.045	ug/l	99
70) Styrene	12.050	104	215469	53.941	ug/l	100
71) Bromoform	12.215	173	40198	48.486	ug/l #	98
73) Isopropylbenzene	12.337	105	400033	51.174	ug/l	99
74) N-amyl acetate	12.154	43	73287	46.663	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.593	83	69902	43.888	ug/l	99
76) 1,2,3-Trichloropropane	12.641	75	45714m	44.007	ug/l	
77) Bromobenzene	12.617	156	83496	48.514	ug/l	98
78) n-propylbenzene	12.684	91	485256	51.062	ug/l	100
79) 2-Chlorotoluene	12.763	91	263810	50.061	ug/l	100
80) 1,3,5-Trimethylbenzene	12.824	105	323183	52.275	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.385	75	21026	45.961	ug/l	99
82) 4-Chlorotoluene	12.867	91	273250	50.754	ug/l	100
83) tert-Butylbenzene	13.086	119	285991	52.521	ug/l	100
84) 1,2,4-Trimethylbenzene	13.129	105	316115	51.921	ug/l	99
85) sec-Butylbenzene	13.263	105	419636	52.009	ug/l	99
86) p-Isopropyltoluene	13.379	119	347736	52.221	ug/l	99
87) 1,3-Dichlorobenzene	13.379	146	168297	49.860	ug/l	99
88) 1,4-Dichlorobenzene	13.458	146	163805	49.617	ug/l	99
89) n-Butylbenzene	13.708	91	338060	52.802	ug/l	100
90) Hexachloroethane	13.970	117	71052	49.656	ug/l	99
91) 1,2-Dichlorobenzene	13.751	146	142266	48.771	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.367	75	8723	43.674	ug/l	98
93) 1,2,4-Trichlorobenzene	15.019	180	77112	53.235	ug/l	99
94) Hexachlorobutadiene	15.123	225	46784	51.557	ug/l	99
95) Naphthalene	15.245	128	133581	50.520	ug/l	100
96) 1,2,3-Trichlorobenzene	15.434	180	64298	52.278	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010224\
Data File : VY016894.D
Acq On : 02 Jan 2024 14:21
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024

Quant Time: Jan 03 00:17:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010224\
Data File : VY016894.D
Acq On : 02 Jan 2024 14:21
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Quant Time: Jan 03 00:17:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:168 Resp: 150750
Ion Ratio Lower Upper
168 100
99 54.1 44.2 66.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#2
Dichlorodifluoromethane
Concen: 47.846 ug/l
RT: 1.905 min Scan# 14
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 85 Resp: 54432
Ion Ratio Lower Upper
85 100
87 32.6 17.1 51.3





#3

Chloromethane

Concen: 48.126 ug/l

RT: 2.119 min Scan# 49

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion: 50 Resp: 85229

Ion Ratio Lower Upper

50 100





52 33.0 26.7 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024

Supervised By :Semsettin Yesilyurt 01/04/2024



#4

Vinyl Chloride

Concen: 48.749 ug/l

RT: 2.253 min Scan# 71

Delta R.T. -0.006 min

Lab File: VY016894.D




Acq: 02 Jan 2024 14:21


Tgt Ion: 62 Resp: 101834

Ion Ratio Lower Upper

62 100


64 32.9 25.7 38.5





#5
Bromomethane
Concen: 48.727 ug/l
RT: 2.661 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 94 Resp: 7292
Ion Ratio Lower Upper
94 100
96 90.7 77.3 115.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#6
Chloroethane
Concen: 48.353 ug/l
RT: 2.802 min Scan# 161
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 64 Resp: 68326
Ion Ratio Lower Upper
64 100
66 32.1 25.1 37.7





#7
Trichlorofluoromethane
Concen: 48.881 ug/l
RT: 3.131 min Scan# 216
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument : MSVOA_Y
ClientSampleId : ICVVY010224




Tgt Ion:101 Resp: 12581
Ion Ratio Lower Upper
101 100
103 61.8 51.9 77.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024




#8
Diethyl Ether
Concen: 47.596 ug/l
RT: 3.533 min Scan# 281
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21



Tgt Ion: 74 Resp: 39150
Ion Ratio Lower Upper
74 100
45 97.3 47.9 143.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.194 ug/l
RT: 3.911 min Scan# 343
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Tgt Ion:101 Resp: 8065
Ion Ratio Lower Upper
101 100
85 45.0 35.8 53.6
151 73.4 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#10
Methyl Iodide
Concen: 51.835 ug/l
RT: 4.100 min Scan# 374
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:142 Resp: 73800
Ion Ratio Lower Upper
142 100
127 43.5 35.1 52.7
141 14.2 11.8 17.6






#11
Tert butyl alcohol
Concen: 221.493 ug/l
RT: 4.954 min Scan# 516
Delta R.T. -0.012 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224






Tgt Ion: 59 Resp: 24419
Ion Ratio Lower Upper
59 100
57 9.7 7.4 11.0


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024




#12
1,1-Dichloroethene
Concen: 49.843 ug/l
RT: 3.881 min Scan# 338
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21



Tgt Ion: 96 Resp: 71569
Ion Ratio Lower Upper
96 100
61 164.0 124.2 186.2
98 63.1 52.6 79.0









#13
Acrolein
Concen: 220.149 ug/l
RT: 3.734 min Scan# 315
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument : MSVOA_Y
ClientSampleId : ICVVY010224

Tgt Ion: 56 Resp: 35924
Ion Ratio Lower Upper
56 100
55 71.1 55.4 83.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#14
Allyl chloride
Concen: 50.866 ug/l
RT: 4.484 min Scan# 437
Delta R.T. -0.013 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 41 Resp: 115466
Ion Ratio Lower Upper
41 100
39 68.4 54.7 82.1
76 39.5 31.8 47.6





#15
Acrylonitrile
Concen: 227.431 ug/l
RT: 5.167 min Scan# 54
Delta R.T. -0.012 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 53 Resp: 8750
Ion Ratio Lower Upper
53 100
52 80.4 65.1 97.7
51 36.3 29.0 43.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#16
Acetone
Concen: 223.763 ug/l
RT: 3.954 min Scan# 350
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 43 Resp: 83103
Ion Ratio Lower Upper
43 100
58 33.1 26.6 39.8





#17
Carbon Disulfide
Concen: 50.929 ug/l
RT: 4.204 min Scan# 391
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 76 Resp: 174734
Ion Ratio Lower Upper
76 100
78 8.7 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#18
Methyl Acetate
Concen: 49.924 ug/l
RT: 4.484 min Scan# 437
Delta R.T. -0.013 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 43 Resp: 79697
Ion Ratio Lower Upper
43 100
74 24.1 19.8 29.6





#19
Methyl tert-butyl Ether
Concen: 48.626 ug/l
RT: 5.228 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 73 Resp: 18165
Ion Ratio Lower Upper
73 100
57 23.1 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#20
Methylene Chloride
Concen: 52.404 ug/l
RT: 4.722 min Scan# 476
Delta R.T. -0.012 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 84 Resp: 99859
Ion Ratio Lower Upper
84 100
49 121.0 95.2 142.8
51 37.8 29.9 44.9
86 61.7 52.2 78.4






#21
trans-1,2-Dichloroethene
Concen: 50.127 ug/l
RT: 5.228 min Scan# 51
Delta R.T. -0.012 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224






Tgt Ion: 96 Resp: 82430
Ion Ratio Lower Upper
96 100
61 132.1 104.6 156.8
98 65.2 50.9 76.3


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024




#22
Diisopropyl ether
Concen: 51.105 ug/l
RT: 6.130 min Scan# 707
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21



Tgt Ion: 45 Resp: 268198
Ion Ratio Lower Upper
45 100
43 50.5 46.8 70.2
87 27.7 22.2 33.2
59 12.2 9.9 14.9









#23
 Vinyl Acetate
 Concen: 238.518 ug/l
 RT: 6.069 min Scan# 697
 Delta R.T. -0.007 min
 Lab File: VY016894.D
 Acq: 02 Jan 2024 14:21

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY010224

Tgt Ion: 43 Resp: 573952
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.6 13.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
 Supervised By :Semsettin Yesilyurt 01/04/2024



#24
 1,1-Dichloroethane
 Concen: 49.219 ug/l
 RT: 6.027 min Scan# 690
 Delta R.T. -0.006 min
 Lab File: VY016894.D
 Acq: 02 Jan 2024 14:21

Tgt Ion: 63 Resp: 155221
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.8 11.4
 100 4.4 2.1 6.2





#25

2-Butanone

Concen: 218.509 ug/l

RT: 6.990 min Scan# 848

Delta R.T. -0.006 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion: 43 Resp: 111728

Ion Ratio Lower Upper

43 100





72 26.8 21.4 32.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024

Supervised By :Semsettin Yesilyurt 01/04/2024



#26

2,2-Dichloropropane

Concen: 50.124 ug/l

RT: 6.990 min Scan# 848

Delta R.T. -0.006 min

Lab File: VY016894.D




Acq: 02 Jan 2024 14:21


Tgt Ion: 77 Resp: 137058

Ion Ratio Lower Upper

77 100


97 23.3 11.7 35.0





#27
cis-1,2-Dichloroethene
Concen: 50.347 ug/l
RT: 6.996 min Scan# 849
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 96 Resp: 97058
Ion Ratio Lower Upper
96 100
61 142.0 0.0 278.8
98 64.1 0.0 127.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#28
Bromochloromethane
Concen: 53.339 ug/l
RT: 7.337 min Scan# 905
Delta R.T. -0.013 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 49 Resp: 63151
Ion Ratio Lower Upper
49 100
129 2.0 0.0 3.8
130 74.4 59.8 89.8





#29

Tetrahydrofuran

Concen: 220.872 ug/l

RT: 7.356 min Scan# 909

Delta R.T. -0.006 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224





Tgt Ion: 42 Resp: 66609

Ion	Ratio	Lower	Upper
42	100		
72	44.8	36.6	54.8
71	41.8	33.8	50.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#30

Chloroform

Concen: 49.334 ug/l

RT: 7.514 min Scan# 934




Delta R.T. -0.007 min


Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Tgt Ion: 83 Resp: 160332


Ion	Ratio	Lower	Upper
83	100		
85	63.7	51.2	76.8





#31
Cyclohexane
Concen: 49.065 ug/l
RT: 7.795 min Scan# 981
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 56 Resp: 12547
Ion Ratio Lower Upper
56 100
69 32.7 27.0 40.6
84 88.9 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#32
1,1,1-Trichloroethane
Concen: 50.223 ug/l
RT: 7.709 min Scan# 966
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 97 Resp: 138395
Ion Ratio Lower Upper
97 100
99 63.9 51.2 76.8
61 43.7 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 45.126 ug/l

RT: 8.148 min Scan# 1039

Delta R.T. -0.007 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion: 65 Resp: 67239

Ion Ratio Lower Upper

65 100





67 55.4 0.0 109.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024

Supervised By :Semsettin Yesilyurt 01/04/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21




Tgt Ion:114 Resp: 250199


Ion Ratio Lower Upper

114 100

63 20.5 0.0 37.4


88 15.0 0.0 29.0





#35
Dibromofluoromethane
Concen: 47.112 ug/l
RT: 7.721 min Scan# 968
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:113 Resp: 75114
Ion Ratio Lower Upper
113 100
111 102.3 83.8 125.6
192 18.1 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#36
1,1-Dichloropropene
Concen: 51.020 ug/l
RT: 7.929 min Scan# 1002
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 75 Resp: 118548
Ion Ratio Lower Upper
75 100
110 34.9 17.4 52.3
77 30.9 24.7 37.1





#37
Ethyl Acetate
Concen: 44.129 ug/l
RT: 7.081 min Scan# 80
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 43 Resp: 4659
Ion Ratio Lower Upper
43 100
61 13.6 11.0 16.6
70 10.1 8.1 12.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#38
Carbon Tetrachloride
Concen: 50.535 ug/l
RT: 7.910 min Scan# 999
Delta R.T. -0.007 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:117 Resp: 119299
Ion Ratio Lower Upper
117 100
119 95.3 76.0 114.0
121 30.2 24.2 36.2






#39
Methylcyclohexane
Concen: 53.008 ug/l
RT: 9.191 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224






Tgt Ion: 83 Resp: 13846
Ion Ratio Lower Upper
83 100
55 75.3 59.7 89.5
98 47.7 38.0 57.0

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#40
Benzene
Concen: 50.333 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 78 Resp: 350830
Ion Ratio Lower Upper
78 100
77 23.5 18.8 28.2





#41
Methacrylonitrile
Concen: 50.362 ug/l
RT: 7.319 min Scan# 90
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:	41	Resp:	2800
Ion Ratio	100	Lower	Upper
39	51.2	41.9	62.9
67	78.1	59.8	89.6
52	30.7	24.8	37.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#42
1,2-Dichloroethane
Concen: 48.093 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:	62	Resp:	88992
Ion Ratio	100	Lower	Upper
98	10.5	0.0	20.2









#43
Isopropyl Acetate
Concen: 45.880 ug/l
RT: 8.282 min Scan# 1060
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument : MSVOA_Y
ClientSampleId : ICVVY010224

Tgt Ion: 43 Resp: 9277
Ion Ratio Lower Upper
43 100
61 31.3 24.8 37.2
87 13.3 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#44
Trichloroethene
Concen: 50.827 ug/l
RT: 8.947 min Scan# 1169
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:130 Resp: 90148
Ion Ratio Lower Upper
130 100
95 98.2 0.0 203.8





#45
1,2-Dichloropropane
Concen: 49.718 ug/l
RT: 9.227 min Scan# 1215
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument : MSVOA_Y
ClientSampleId : ICVVY010224



Tgt Ion: 63 Resp: 8927
Ion Ratio Lower Upper
63 100
65 30.4 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#46
Dibromomethane
Concen: 48.408 ug/l
RT: 9.313 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 93 Resp: 44332
Ion Ratio Lower Upper
93 100
95 84.8 66.6 99.8
174 90.8 73.8 110.6





#47
Bromodichloromethane
Concen: 49.151 ug/l
RT: 9.508 min Scan# 1261
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVY010224



Tgt Ion: 83 Resp: 120872
Ion Ratio Lower Upper
83 100
85 64.0 53.2 79.8
127 7.9 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#48
Methyl methacrylate
Concen: 49.314 ug/l
RT: 9.300 min Scan# 1227
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 41 Resp: 53814
Ion Ratio Lower Upper
41 100
69 89.2 72.1 108.1
39 51.8 42.8 64.2





#49
1,4-Dioxane
Concen: 928.874 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 88 Resp: 8808
Ion Ratio Lower Upper
88 100
43 39.0 28.2 42.4
58 66.5 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#50
Toluene-d8
Concen: 48.097 ug/l
RT: 10.191 min Scan# 1373
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 98 Resp: 264060
Ion Ratio Lower Upper
98 100
100 65.8 52.3 78.5









#51
4-Methyl-2-Pentanone
Concen: 228.923 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Tgt Ion: 43 Resp: 24617
Ion Ratio Lower Upper
43 100
58 40.8 32.2 48.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#52
Toluene
Concen: 51.072 ug/l
RT: 10.252 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 92 Resp: 214834
Ion Ratio Lower Upper
92 100
91 174.3 137.6 206.4





#53
t-1,3-Dichloropropene
Concen: 50.038 ug/l
RT: 10.471 min Scan# 1419
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 75 Resp: 110224
Ion Ratio Lower Upper
75 100
77 31.0 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#54
cis-1,3-Dichloropropene
Concen: 50.656 ug/l
RT: 9.935 min Scan# 1331
Delta R.T. -0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 75 Resp: 134005
Ion Ratio Lower Upper
75 100
77 32.5 24.6 36.8
39 45.0 33.6 50.4





#55
1,1,2-Trichloroethane
Concen: 47.435 ug/l
RT: 10.654 min Scan# 1449
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument : MSVOA_Y
ClientSampleId : ICVVY010224



Tgt Ion:	97	Resp:	6346
Ion	Ratio	Lower	Upper
97	100		
83	87.7	69.4	104.2
85	57.9	43.9	65.9
99	65.8	48.6	73.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#56
Ethyl methacrylate
Concen: 49.165 ug/l
RT: 10.520 min Scan# 1427
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:	69	Resp:	55715
Ion	Ratio	Lower	Upper
69	100		
41	63.7	49.2	73.8
39	34.3	26.8	40.2





#57
1,3-Dichloropropane
Concen: 48.814 ug/l
RT: 10.800 min Scan# 1473
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 76 Resp: 109039
Ion Ratio Lower Upper
76 100
78 32.4 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#58
2-Chloroethyl Vinyl ether
Concen: 235.967 ug/l
RT: 9.788 min Scan# 1307
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 63 Resp: 167516
Ion Ratio Lower Upper
63 100
106 27.7 22.4 33.6





#59
2-Hexanone
Concen: 218.016 ug/l
RT: 10.843 min Scan# 1480
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 43 Resp: 17263
Ion Ratio Lower Upper
43 100
58 55.9 28.1 84.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#60
Dibromochloromethane
Concen: 48.766 ug/l
RT: 10.995 min Scan# 1505
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:129 Resp: 75242
Ion Ratio Lower Upper
129 100
127 77.7 39.2 117.6





#61
1,2-Dibromoethane
Concen: 49.190 ug/l
RT: 11.099 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224




Tgt Ion:107 Resp: 57180
Ion Ratio Lower Upper
107 100
109 91.2 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024




#62
4-Bromofluorobenzene
Concen: 47.184 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Tgt Ion: 95 Resp: 87401
Ion Ratio Lower Upper
95 100
174 80.8 0.0 160.0
176 78.0 0.0 151.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.501 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:117 Resp: 212664
Ion Ratio Lower Upper
117 100
82 56.6 43.8 65.8
119 32.7 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#64
Tetrachloroethene
Concen: 50.245 ug/l
RT: 10.727 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:164 Resp: 69275
Ion Ratio Lower Upper
164 100
166 126.0 98.8 148.2
129 99.4 78.9 118.3
131 92.7 72.5 108.7





#65
Chlorobenzene
Concen: 51.207 ug/l
RT: 11.526 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:112 Resp: 226858
Ion Ratio Lower Upper
112 100
114 32.1 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 50.579 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:131 Resp: 85180
Ion Ratio Lower Upper
131 100
133 94.5 47.0 141.0
119 68.1 33.3 99.9





#67
Ethyl Benzene
Concen: 52.828 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 91 Resp: 421349
Ion Ratio Lower Upper
91 100
106 30.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#68
m/p-Xylenes
Concen: 106.392 ug/l
RT: 11.709 min Scan# 1622
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:106 Resp: 311476
Ion Ratio Lower Upper
106 100
91 202.1 160.7 241.1









#69
o-Xylene
Concen: 53.045 ug/l
RT: 12.038 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Tgt Ion:106 Resp: 14558
Ion Ratio Lower Upper
106 100
91 214.1 106.5 319.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#70
Styrene
Concen: 53.941 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:104 Resp: 215469
Ion Ratio Lower Upper
104 100
78 49.0 39.2 58.8
103 55.1 43.8 65.6





#71

Bromoform

Concen: 48.486 ug/l

RT: 12.215 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion:173 Resp: 40198

Ion Ratio Lower Upper

173 100





175 48.4 23.5 70.6

254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.434 min Scan# 1905

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21




Tgt Ion:152 Resp: 95726


Ion Ratio Lower Upper

152 100

115 58.0 30.0 90.0

150 170.6 0.0 344.6





#73

Isopropylbenzene

Concen: 51.174 ug/l

RT: 12.337 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion:105 Resp: 40003

Ion Ratio Lower Upper

105 100





120 26.5 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024

Supervised By :Semsettin Yesilyurt 01/04/2024



#74

N-ethyl acetate

Concen: 46.663 ug/l

RT: 12.154 min Scan# 1695

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Tgt Ion: 43 Resp: 73287




Ion Ratio Lower Upper


43 100

70 45.0 36.6 55.0

55 26.7 22.2 33.4


61 24.9 20.2 30.4





#75
1,1,2,2-Tetrachloroethane
Concen: 43.888 ug/l
RT: 12.593 min Scan# 1767
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 83 Resp: 6990
Ion Ratio Lower Upper
83 100
131 9.3 4.9 14.6
85 64.9 32.6 98.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#76
1,2,3-Trichloropropane
Concen: 44.007 ug/l m
RT: 12.641 min Scan# 1775
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion: 75 Resp: 45714
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 48.514 ug/l

RT: 12.617 min Scan# 1771

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :





ICVVY010224

Tgt Ion:	156	Resp:	83490
Ion Ratio	Lower	Upper	
156	100		
77	190.0	97.3	291.8
158	96.5	48.4	145.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#78

n-propylbenzene

Concen: 51.062 ug/l




RT: 12.684 min Scan# 1782


Delta R.T. 0.006 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Tgt Ion:	91	Resp:	485256
Ion Ratio	Lower	Upper	
91	100		
120	22.2	11.0	33.0









#79
2-Chlorotoluene
Concen: 50.061 ug/l
RT: 12.763 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224

Tgt Ion: 91 Resp: 263810
Ion Ratio Lower Upper
91 100
126 34.0 17.0 51.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#80
1,3,5-Trimethylbenzene
Concen: 52.275 ug/l
RT: 12.824 min Scan# 1805
Delta R.T. 0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:105 Resp: 323183
Ion Ratio Lower Upper
105 100
120 48.7 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 45.961 ug/l
RT: 12.385 min Scan# 1733
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion: 75 Resp: 21020
Ion Ratio Lower Upper
75 100
53 91.2 74.0 111.0
89 45.8 36.5 54.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#82
4-Chlorotoluene
Concen: 50.754 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. 0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion: 91 Resp: 273250
Ion Ratio Lower Upper
91 100
126 34.3 17.1 51.2





#83
tert-Butylbenzene
Concen: 52.521 ug/l
RT: 13.086 min Scan# 1848
Delta R.T. -0.001 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument : MSVOA_Y
ClientSampleId : ICVVY010224



Tgt Ion:119 Resp: 28599
Ion Ratio Lower Upper
119 100
91 63.0 31.5 94.5
134 26.1 13.1 39.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#84
1,2,4-Trimethylbenzene
Concen: 51.921 ug/l
RT: 13.129 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21


Tgt Ion:105 Resp: 316115
Ion Ratio Lower Upper
105 100
120 45.1 22.1 66.3





#85
 sec-Butylbenzene
 Concen: 52.009 ug/l
 RT: 13.263 min Scan# 1877
 Delta R.T. 0.000 min
 Lab File: VY016894.D
 Acq: 02 Jan 2024 14:21




Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY010224




Tgt Ion:105 Resp: 419630
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.7 28.9

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
 Supervised By :Semsettin Yesilyurt 01/04/2024




#86
 p-Isopropyltoluene
 Concen: 52.221 ug/l
 RT: 13.379 min Scan# 1896
 Delta R.T. 0.000 min
 Lab File: VY016894.D
 Acq: 02 Jan 2024 14:21




Tgt Ion:119 Resp: 347736
 Ion Ratio Lower Upper
 119 100
 134 26.6 13.1 39.1
 91 23.1 11.7 35.0





#87
1,3-Dichlorobenzene
Concen: 49.860 ug/l
RT: 13.379 min Scan# 1895
Delta R.T. 0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:146 Resp: 168297
Ion Ratio Lower Upper
146 100
111 40.2 20.2 60.5
148 63.2 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#88
1,4-Dichlorobenzene
Concen: 49.617 ug/l
RT: 13.458 min Scan# 1909
Delta R.T. 0.006 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:146 Resp: 163805
Ion Ratio Lower Upper
146 100
111 38.3 19.3 57.9
148 62.7 31.6 94.7





#89

n-Butylbenzene

Concen: 52.802 ug/l

RT: 13.708 min Scan# 1949

Delta R.T. 0.006 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY010224

Tgt Ion: 91 Resp: 338060

Ion Ratio Lower Upper

91 100

92 52.7 26.4 79.2





134 24.4 12.3 36.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024

Supervised By :Semsettin Yesilyurt 01/04/2024



#90

Hexachloroethane

Concen: 49.656 ug/l

RT: 13.970 min Scan# 1993

Delta R.T. -0.001 min

Lab File: VY016894.D




Acq: 02 Jan 2024 14:21


Tgt Ion:117 Resp: 71052

Ion Ratio Lower Upper

117 100


201 67.1 33.4 100.1





#91
1,2-Dichlorobenzene
Concen: 48.771 ug/l
RT: 13.751 min Scan# 1957
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224




Tgt Ion:146 Resp: 142260
Ion Ratio Lower Upper
146 100
111 41.4 20.5 61.6
148 64.5 31.4 94.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 43.674 ug/l
RT: 14.367 min Scan# 2058
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21



Tgt Ion: 75 Resp: 8723
Ion Ratio Lower Upper
75 100
155 85.1 42.1 126.3
157 108.6 55.8 167.4





#93

1,2,4-Trichlorobenzene

Concen: 53.235 ug/l

RT: 15.019 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21

Instrument :

MSVOA_Y

ClientSampleId :





ICVVY010224

Tgt Ion:180	Resp:	7711
Ion Ratio	Lower	Upper
180	100	
182	95.0	47.9 143.7
145	30.8	15.9 47.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#94

Hexachlorobutadiene

Concen: 51.557 ug/l




RT: 15.123 min Scan# 2182


Delta R.T. -0.000 min

Lab File: VY016894.D

Acq: 02 Jan 2024 14:21


Tgt Ion:225	Resp:	46784
Ion Ratio	Lower	Upper
225	100	
223	63.9	31.4 94.3
227	63.6	32.3 96.8





#95
Naphthalene
Concen: 50.520 ug/l
RT: 15.245 min Scan# 2202
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010224



Tgt Ion:128 Resp: 13358
Ion Ratio Lower Upper
128 100
127 12.9 10.2 15.4
129 10.7 8.6 12.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/04/2024
Supervised By :Semsettin Yesilyurt 01/04/2024



#96
1,2,3-Trichlorobenzene
Concen: 52.278 ug/l
RT: 15.434 min Scan# 2233
Delta R.T. -0.000 min
Lab File: VY016894.D
Acq: 02 Jan 2024 14:21

Tgt Ion:180 Resp: 64298
Ion Ratio Lower Upper
180 100
182 94.9 46.3 138.8
145 31.7 16.1 48.2

