

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010324\
 Data File : VY016905.D
 Acq On : 03 Jan 2024 10:51
 Operator : SY/MD
 Sample : VY0103SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0103SBL01

Quant Time: Jan 04 00:29:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	117016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.710	114	250283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	239685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	89596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	62019	53.625	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 107.240%			
35) Dibromofluoromethane	7.734	113	63390	39.746	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 79.500%			
50) Toluene-d8	10.197	98	230334	41.940	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 83.880%			
62) 4-Bromofluorobenzene	12.495	95	73211	39.511	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 79.020%			
Target Compounds						
20) Methylene Chloride	4.747	84	19048	6.403	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010324\
Data File : VY016905.D
Acq On : 03 Jan 2024 10:51
Operator : SY/MD
Sample : VY0103SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0103SBL01

Quant Time: Jan 04 00:29:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

