

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010524\
 Data File : VY016921.D
 Acq On : 05 Jan 2024 14:39
 Operator : SY/MD
 Sample : P1040-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Jan 05 22:36:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	116892	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	246389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	237516	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	92476	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	59012	51.079	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.160%
35) Dibromofluoromethane	7.740	113	62271	39.661	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.320%
50) Toluene-d8	10.197	98	224042	41.439	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.880%
62) 4-Bromofluorobenzene	12.496	95	72083	39.517	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010524\
Data File : VY016921.D
Acq On : 05 Jan 2024 14:39
Operator : SY/MD
Sample : P1040-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Jan 05 22:36:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

