

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010622\
 Data File : VY007028.D
 Acq On : 06 Jan 2022 13:32
 Operator : SY/MD
 Sample : M5342-05RE
 Misc : 2.98g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8-(0-2)RE

Quant Time: Jan 06 15:36:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 06 04:51:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	134479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	239372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	178504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	55172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	78615	49.524	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.040%
35) Dibromofluoromethane	7.722	113	80897	51.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.820%
50) Toluene-d8	10.185	98	270017	46.186	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.380%
62) 4-Bromofluorobenzene	12.483	95	71730	32.579	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.160%
Target Compounds						
					Qvalue	
44) Trichloroethene	8.947	130	6344	3.183	ug/l	98
64) Tetrachloroethene	10.721	164	7736	5.712	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010622\
Data File : VY007028.D
Acq On : 06 Jan 2022 13:32
Operator : SY/MD
Sample : M5342-05RE
Misc : 2.98g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8-(0-2)RE

Quant Time: Jan 06 15:36:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 06 04:51:57 2022
Response via : Initial Calibration

