

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010622\
 Data File : VY007037.D
 Acq On : 06 Jan 2022 17:05
 Operator : SY/MD
 Sample : N1008-06RE
 Misc : 4.78g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOM-2022-3BRE

Quant Time: Jan 07 11:48:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 06 04:51:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	66054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	123264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	115274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	51714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	44462	57.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.040%
35) Dibromofluoromethane	7.722	113	41087	50.708	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.420%
50) Toluene-d8	10.185	98	150970	50.148	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.300%
62) 4-Bromofluorobenzene	12.483	95	56234	49.599	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.200%
Target Compounds						
16) Acetone	3.942	43	1492	8.685	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010622\
Data File : VY007037.D
Acq On : 06 Jan 2022 17:05
Operator : SY/MD
Sample : N1008-06RE
Misc : 4.78g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOM-2022-3BRE

Quant Time: Jan 07 11:48:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 06 04:51:57 2022
Response via : Initial Calibration

