


Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 3.669\text{e-}001 * \text{Amt} + 3.046\text{e-}002$$


Coef of Det (r^2) = 0.999234 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Chloromethane

Response Ratio



$$\text{Response} = 3.400\text{e-}001 * \text{Amt} + 2.331\text{e-}002$$


Coef of Det (r^2) = 0.998219 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Bromomethane

Response Ratio



$$\text{Response} = 2.913\text{e-}001 * \text{Amt} + 2.676\text{e-}002$$


Coef of Det (r^2) = 0.999502 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Tert butyl alcohol

Response Ratio




$$\text{Response} = 5.689\text{e-}002 * \text{Amt} + 4.352\text{e-}002$$

Coef of Det (r^2) = 0.997255 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023


Acrolein



$\text{Response} = 2.485\text{e-}002 * \text{Amt} + 2.039\text{e-}002$
 Coef of Det (r^2) = 0.993018 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Carbon Disulfide

Response Ratio



$$\text{Response} = 1.067\text{e}+000 * \text{Amt} + 1.113\text{e}-001$$


Coef of Det (r^2) = 0.997539 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Methylene Chloride

Response Ratio



$$\text{Response} = 4.926\text{e-}001 * \text{Amt} + 6.612\text{e-}002$$


Coef of Det (r^2) = 0.999749 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M

Calibration Table Last Updated: Sat Jan 07 01:46:48 2023

Cyclohexane

Response Ratio



$\text{Response} = 6.598\text{e-}001 * \text{Amt} + 4.979\text{e-}002$
 Coef of Det (r^2) = 0.999073 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Calibration Table Last Updated: Sat Jan 07 01:46:48 2023