

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010623\
 Data File : VY012091.D
 Acq On : 06 Jan 2023 14:39
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY010623

Quant Time: Jan 07 01:48:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.448	0.396	11.6	91	0.00
3 P	Chloromethane	0.411	0.365	11.2	94	0.00
4 C	Vinyl Chloride	0.486	0.431	11.3#	91	0.00
5 T	Bromomethane	0.372	0.346	7.0	101	0.00
6 T	Chloroethane	0.322	0.296	8.1	94	0.00
7 T	Trichlorofluoromethane	0.968	0.893	7.7	91	0.00
8 T	Diethyl Ether	0.296	0.290	2.0	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.498	0.479	3.8	92	0.00
10 T	Methyl Iodide	0.676	0.605	10.5	85	0.00
11 T	Tert butyl alcohol	0.083	0.063	24.1	94	0.00
12 CM	1,1-Dichloroethene	0.489	0.462	5.5#	93	0.00
13 T	Acrolein	0.036	0.028	22.2	97	0.00
14 T	Allyl chloride	0.659	0.654	0.8	96	0.00
15 T	Acrylonitrile	0.139	0.145	-4.3	100	0.00
16 T	Acetone	0.132	0.136	-3.0	101	0.00
17 T	Carbon Disulfide	1.378	1.147	16.8	93	0.00
18 T	Methyl Acetate	0.380	0.371	2.4	99	0.00
19 T	Methyl tert-butyl Ether	1.438	1.468	-2.1	95	0.00
20 T	Methylene Chloride	0.688	0.604	12.2	98	0.00
21 T	trans-1,2-Dichloroethene	0.541	0.507	6.3	92	0.00
22 T	Diisopropyl ether	1.355	1.414	-4.4	97	0.00
23 T	Vinyl Acetate	0.820	0.870	-6.1	98	0.00
24 P	1,1-Dichloroethane	0.901	0.899	0.2	95	0.00
25 T	2-Butanone	0.174	0.178	-2.3	100	0.00
26 T	2,2-Dichloropropane	0.972	0.921	5.2	91	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.614	0.6	93	0.00
28 T	Bromochloromethane	0.228	0.233	-2.2	100	0.00
29 T	Tetrahydrofuran	0.102	0.108	-5.9	100	0.00
30 C	Chloroform	1.028	1.032	-0.4#	94	0.00
31 T	Cyclohexane	0.808	0.706	12.6	91	0.00
32 T	1,1,1-Trichloroethane	0.998	0.980	1.8	92	0.00
33 S	1,2-Dichloroethane-d4	0.562	0.526	6.4	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.329	0.314	4.6	93	0.00
36 T	1,1-Dichloropropene	0.499	0.457	8.4	92	0.00
37 T	Ethyl Acetate	0.267	0.250	6.4	99	0.00
38 T	Carbon Tetrachloride	0.633	0.598	5.5	91	0.00
39 T	Methylcyclohexane	0.596	0.550	7.7	91	0.00
40 TM	Benzene	1.453	1.364	6.1	93	0.00
41 T	Methacrylonitrile	0.145	0.131	9.7	82	-0.02
42 TM	1,2-Dichloroethane	0.460	0.447	2.8	93	0.00
43 T	Isopropyl Acetate	0.447	0.458	-2.5	98	0.00
44 TM	Trichloroethene	0.424	0.403	5.0	93	0.00
45 C	1,2-Dichloropropane	0.324	0.315	2.8#	94	0.00
46 T	Dibromomethane	0.204	0.202	1.0	95	0.00
47 T	Bromodichloromethane	0.518	0.519	-0.2	95	0.00
48 T	Methyl methacrylate	0.200	0.205	-2.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010623\
 Data File : VY012091.D
 Acq On : 06 Jan 2023 14:39
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY010623

Quant Time: Jan 07 01:48:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	103	0.00
50 S	Toluene-d8	1.297	1.182	8.9	92	0.00
51 T	4-Methyl-2-Pentanone	0.238	0.249	-4.6	99	0.00
52 CM	Toluene	0.962	0.908	5.6#	93	0.00
53 T	t-1,3-Dichloropropene	0.539	0.537	0.4	95	0.00
54 T	cis-1,3-Dichloropropene	0.583	0.579	0.7	94	0.00
55 T	1,1,2-Trichloroethane	0.294	0.298	-1.4	95	0.00
56 T	Ethyl methacrylate	0.386	0.398	-3.1	96	0.00
57 T	1,3-Dichloropropane	0.486	0.490	-0.8	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.138	2.1	95	0.00
59 T	2-Hexanone	0.167	0.177	-6.0	98	0.00
60 T	Dibromochloromethane	0.388	0.390	-0.5	95	0.00
61 T	1,2-Dibromoethane	0.282	0.276	2.1	96	0.00
62 S	4-Bromofluorobenzene	0.453	0.430	5.1	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.490	0.453	7.6	91	0.00
65 PM	Chlorobenzene	1.142	1.096	4.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.441	0.434	1.6	94	0.00
67 C	Ethyl Benzene	2.041	1.952	4.4#	92	0.00
68 T	m/p-Xylenes	0.801	0.768	4.1	93	0.00
69 T	o-Xylene	0.762	0.738	3.1	93	0.00
70 T	Styrene	1.295	1.260	2.7	93	0.00
71 P	Bromoform	0.280	0.282	-0.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.956	3.760	5.0	93	0.00
74 T	N-amyl acetate	0.817	0.848	-3.8	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.666	0.667	-0.2	99	0.00
76 T	1,2,3-Trichloropropane	0.571	0.545	4.6	102	0.00
77 T	Bromobenzene	0.944	0.893	5.4	95	0.00
78 T	n-propylbenzene	4.619	4.407	4.6	94	0.00
79 T	2-Chlorotoluene	2.642	2.506	5.1	95	0.00
80 T	1,3,5-Trimethylbenzene	3.393	3.284	3.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.257	0.254	1.2	97	0.00
82 T	4-Chlorotoluene	2.763	2.615	5.4	94	0.00
83 T	tert-Butylbenzene	2.973	2.860	3.8	93	0.00
84 T	1,2,4-Trimethylbenzene	3.359	3.241	3.5	95	0.00
85 T	sec-Butylbenzene	4.273	4.121	3.6	94	0.00
86 T	p-Isopropyltoluene	3.702	3.592	3.0	95	0.00
87 T	1,3-Dichlorobenzene	1.892	1.796	5.1	95	0.00
88 T	1,4-Dichlorobenzene	1.890	1.793	5.1	96	0.00
89 T	n-Butylbenzene	3.208	3.165	1.3	96	0.00
90 T	Hexachloroethane	0.687	0.663	3.5	95	0.00
91 T	1,2-Dichlorobenzene	1.685	1.625	3.6	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.132	0.133	-0.8	102	0.00
93 T	1,2,4-Trichlorobenzene	1.051	1.061	-1.0	101	0.00
94 T	Hexachlorobutadiene	0.634	0.631	0.5	99	0.00
95 T	Naphthalene	2.079	2.165	-4.1	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010623\
Data File : VY012091.D
Acq On : 06 Jan 2023 14:39
Operator : KP/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY010623

Quant Time: Jan 07 01:48:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.913	0.921	-0.9	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6