

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
 Data File : VY007058.D
 Acq On : 07 Jan 2022 17:15
 Operator : SY/MD
 Sample : N1065-16
 Misc : 6.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AMI-4

Quant Time: Jan 08 04:11:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 06 04:51:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	123865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	235678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	221421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	90541	61.924	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.840%	
35) Dibromofluoromethane	7.722	113	79137	51.082	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.160%	
50) Toluene-d8	10.185	98	287310	49.915	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.820%	
62) 4-Bromofluorobenzene	12.483	95	108921	50.246	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.500%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.123	105	15675	2.191	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
Data File : VY007058.D
Acq On : 07 Jan 2022 17:15
Operator : SY/MD
Sample : N1065-16
Misc : 6.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AMI-4

Quant Time: Jan 08 04:11:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 06 04:51:57 2022
Response via : Initial Calibration

