

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
 Data File : VY007060.D
 Acq On : 07 Jan 2022 18:02
 Operator : SY/MD
 Sample : N1065-24
 Misc : 2.90g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AMI-6

Quant Time: Jan 08 04:12:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 06 04:51:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	83989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	155098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	142565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	65726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	57930	58.431	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.860%
35) Dibromofluoromethane	7.722	113	50411	49.446	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.900%
50) Toluene-d8	10.184	98	186560	49.250	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.500%
62) 4-Bromofluorobenzene	12.483	95	70807	49.634	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.260%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	6188	28.330	ug/l	94
18) Methyl Acetate	4.478	43	20021	21.365	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
Data File : VY007060.D
Acq On : 07 Jan 2022 18:02
Operator : SY/MD
Sample : N1065-24
Misc : 2.90g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AMI-6

Quant Time: Jan 08 04:12:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 06 04:51:57 2022
Response via : Initial Calibration

