

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
 Data File : VY007064.D
 Acq On : 07 Jan 2022 19:37
 Operator : SY/MD
 Sample : N1065-40
 Misc : 4.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AMI-10

Quant Time: Jan 08 04:13:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 06 04:51:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	27751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	49210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	40520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	16692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	16899	51.588	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.180%
35) Dibromofluoromethane	7.728	113	16129	49.861	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.720%
50) Toluene-d8	10.185	98	57529	47.866	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.740%
62) 4-Bromofluorobenzene	12.483	95	18400	40.651	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.300%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	3800	52.654	ug/l	97
18) Methyl Acetate	4.485	43	34455	111.279	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010722\
Data File : VY007064.D
Acq On : 07 Jan 2022 19:37
Operator : SY/MD
Sample : N1065-40
Misc : 4.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AMI-10

Quant Time: Jan 08 04:13:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010522S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 06 04:51:57 2022
Response via : Initial Calibration

