

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
 Data File : VY007087.D
 Acq On : 08 Jan 2022 18:38
 Operator : SY/MD
 Sample : N1071-01
 Misc : 6.37g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FM-E6-01-4.5

Quant Time: Jan 10 03:42:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	131727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	232274	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	210846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	88117	55.916	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.840%
35) Dibromofluoromethane	7.722	113	76316	49.298	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.600%
50) Toluene-d8	10.185	98	281416	49.215	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.420%
62) 4-Bromofluorobenzene	12.483	95	98960	47.986	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
Data File : VY007087.D
Acq On : 08 Jan 2022 18:38
Operator : SY/MD
Sample : N1071-01
Misc : 6.37g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FM-E6-01-4.5

Quant Time: Jan 10 03:42:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

