

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
 Data File : VY007090.D
 Acq On : 08 Jan 2022 19:46
 Operator : SY/MD
 Sample : N1065-03
 Misc : 1.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AMI-1

Quant Time: Jan 10 03:42:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	125782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	224736	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	203448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	82008	54.499	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.000%	
35) Dibromofluoromethane	7.722	113	71662	47.844	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 95.680%	
50) Toluene-d8	10.185	98	272299	49.217	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.440%	
62) 4-Bromofluorobenzene	12.483	95	96682	48.454	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.900%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	6186	21.911	ug/l	92
18) Methyl Acetate	4.491	43	5166	3.989	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
Data File : VY007090.D
Acq On : 08 Jan 2022 19:46
Operator : SY/MD
Sample : N1065-03
Misc : 1.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AMI-1

Quant Time: Jan 10 03:42:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

