

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
 Data File : VY007091.D
 Acq On : 08 Jan 2022 20:09
 Operator : SY/MD
 Sample : N1065-08
 Misc : 6.87g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AMI-2

Quant Time: Jan 10 03:43:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	97034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	173465	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	154971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	64685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	64911	55.917	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.840%
35) Dibromofluoromethane	7.722	113	56059	48.490	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.980%
50) Toluene-d8	10.185	98	210209	49.225	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.483	95	73615	47.798	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.600%
Target Compounds						
16) Acetone	3.954	43	11811	54.230	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010822\
Data File : VY007091.D
Acq On : 08 Jan 2022 20:09
Operator : SY/MD
Sample : N1065-08
Misc : 6.87g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AMI-2

Quant Time: Jan 10 03:43:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

