

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010824\
 Data File : VY016935.D
 Acq On : 08 Jan 2024 14:28
 Operator : SY/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Jan 09 06:36:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	116695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	249524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	242177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	94676	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	57273	49.657	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.320%
35) Dibromofluoromethane	7.734	113	61138	38.450	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.900%
50) Toluene-d8	10.191	98	229158	41.852	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	83.700%
62) 4-Bromofluorobenzene	12.495	95	75564	40.904	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010824\
Data File : VY016935.D
Acq On : 08 Jan 2024 14:28
Operator : SY/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Jan 09 06:36:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

