



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010824\
Data File : VY016926.D
Acq On : 08 Jan 2024 08:44
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Title : SW846 8260
Last Update : Tue Jan 02 14:25:32 2024



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	15888	PASS
75	95	30	60	53.3	41867	PASS
95	95	100	100	100.0	78581	PASS
96	95	5	9	6.9	5442	PASS
173	174	0.00	2	1.1	678	PASS
174	95	50	100	76.6	60197	PASS
175	174	5	9	7.8	4694	PASS
176	174	95	101	96.8	58267	PASS
177	176	5	9	6.8	3964	PASS