

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011024\
 Data File : VY016970.D
 Acq On : 10 Jan 2024 14:03
 Operator : SY/MD
 Sample : P1085-07
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-4

Quant Time: Jan 11 00:12:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	93411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	175689	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	145667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	49753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	30905	33.475	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	66.940%
35) Dibromofluoromethane	7.728	113	36119	32.262	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	64.520%
50) Toluene-d8	10.191	98	151935	39.410	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	78.820%
62) 4-Bromofluorobenzene	12.495	95	38881	29.892	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	59.780%
Target Compounds						
17) Carbon Disulfide	4.204	76	2155	1.014	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011024\
Data File : VY016970.D
Acq On : 10 Jan 2024 14:03
Operator : SY/MD
Sample : P1085-07
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-4

Quant Time: Jan 11 00:12:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

