

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
 Data File : VY007116.D
 Acq On : 11 Jan 2022 15:37
 Operator : SY/MD
 Sample : N1099-02
 Misc : 5.92g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-02

Quant Time: Jan 12 05:09:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	126433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	229065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	215605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	102734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	87979	58.166	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.340%
35) Dibromofluoromethane	7.728	113	74588	48.857	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.720%
50) Toluene-d8	10.185	98	280744	49.785	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.483	95	106492	52.362	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.720%
Target Compounds						
16) Acetone	3.967	43	15754	55.514	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
Data File : VY007116.D
Acq On : 11 Jan 2022 15:37
Operator : SY/MD
Sample : N1099-02
Misc : 5.92g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-02

Quant Time: Jan 12 05:09:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

