

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
 Data File : VY007120.D
 Acq On : 11 Jan 2022 17:16
 Operator : SY/MD
 Sample : N1099-06
 Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-06

Quant Time: Jan 12 05:09:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	232461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	219738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81723	54.781	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.560%	
35) Dibromofluoromethane	7.722	113	77009	49.706	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.420%	
50) Toluene-d8	10.185	98	283512	49.541	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.080%	
62) 4-Bromofluorobenzene	12.483	95	105280	51.009	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.020%	
Target Compounds						
16) Acetone	3.954	43	4964	17.736	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
Data File : VY007120.D
Acq On : 11 Jan 2022 17:16
Operator : SY/MD
Sample : N1099-06
Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-06

Quant Time: Jan 12 05:09:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

