

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
 Data File : VY007123.D
 Acq On : 11 Jan 2022 18:50
 Operator : SY/MD
 Sample : N1099-09
 Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Jan 12 05:10:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.79
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.70
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.50
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
Data File : VY007123.D
Acq On : 11 Jan 2022 18:50
Operator : SY/MD
Sample : N1099-09
Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Jan 12 05:10:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration

