

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
 Data File : VY007120.D
 Acq On : 11 Jan 2022 17:16
 Operator : SY/MD
 Sample : N1099-06
 Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-06

Quant Time: Jan 12 05:09:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 10 03:29:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.795	168	124698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	232461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	219738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	95804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81723	54.781	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.560%	
35) Dibromofluoromethane	7.722	113	77009	49.706	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.420%	
50) Toluene-d8	10.185	98	283512	49.541	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.080%	
62) 4-Bromofluorobenzene	12.483	95	105280	51.009	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.020%	
Target Compounds						Qvalue
16) Acetone	3.954	43	4964	17.736	ug/l	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011122\
Data File : VY007120.D
Acq On : 11 Jan 2022 17:16
Operator : SY/MD
Sample : N1099-06
Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-06

Quant Time: Jan 12 05:09:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010822S.M
Quant Title : SW846 8260
QLast Update : Mon Jan 10 03:29:44 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16




Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-06


Tgt Ion:168 Resp: 124698
Ion Ratio Lower Upper
168 100
99 65.5 49.1 73.7



#16
Acetone
Concen: 17.736 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16

Tgt Ion: 43 Resp: 4964
Ion Ratio Lower Upper
43 100
58 25.8 23.8 35.6





#33

1,2-Dichloroethane-d4

Concen: 54.781 ug/l

RT: 8.149 min Scan# 1038

Delta R.T. 0.000 min

Lab File: VY007120.D

Acq: 11 Jan 2022 17:16

Instrument :

MSVOA_Y

ClientSampleId :





VACLOT-BACKFILL-TILCON-06

Tgt Ion: 65 Resp: 81723

Ion Ratio Lower Upper

65 100

67 50.9 0.0 105.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VY007120.D

Acq: 11 Jan 2022 17:16




Tgt Ion: 114 Resp: 232461


Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.2

88 15.2 0.0 31.0









#35
Dibromofluoromethane
Concen: 49.706 ug/l
RT: 7.722 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16




Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-06


Tgt Ion:113 Resp: 77009
Ion Ratio Lower Upper
113 100
111 103.2 82.2 123.4
192 17.5 14.2 21.4



#50
Toluene-d8
Concen: 49.541 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16

Tgt Ion: 98 Resp: 283512
Ion Ratio Lower Upper
98 100
100 65.4 52.2 78.2









#62
4-Bromofluorobenzene
Concen: 51.009 ug/l
RT: 12.483 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16




Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-06


Tgt Ion: 95 Resp: 105280
Ion Ratio Lower Upper
95 100
174 73.4 0.0 148.4
176 72.1 0.0 146.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY007120.D
Acq: 11 Jan 2022 17:16


Tgt Ion: 117 Resp: 219738
Ion Ratio Lower Upper
117 100
82 57.0 45.4 68.0
119 32.1 26.4 39.6





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.428 min Scan# 1904
 Delta R.T. 0.000 min
 Lab File: VY007120.D
 Acq: 11 Jan 2022 17:16

Instrument : MSVOA_Y
 ClientSampleId : VACLOT-BACKFILL-TILCON-06



Tgt Ion:152 Resp: 95804

Ion	Ratio	Lower	Upper
152	100		
115	63.3	31.4	94.0
150	157.4	0.0	348.4

